

CERTIFICATE OF COMPLIANCE

Certificate Number 20180816-E486483
Report Reference E486483-20170127
Issue Date 2018-AUGUST-16

Issued to: mPower Electronics
3046 Scott Blvd
Santa Clara CA 95054

**This is to certify that
representative samples of**

**GAS AND VAPOR DETECTION EQUIPMENT FOR USE IN
HAZARDOUS LOCATIONS**

Single gas detector, Model MP100, MP101, MP102 and
MP103, intrinsically safe when used with mPower battery
Part No. 1.17.02.0002 or EVE Part No. ER14505.

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety:

UL 913, Intrinsically Safe Apparatus and Associated
Apparatus for Use in Class I, II, and III, Division 1,
Hazardous (Classified) Locations
UL 508 and CSA C22.2 No. 14, Standard for Industrial
Control Equipment
CSA-C22.2 No. 157-92, Intrinsically Safe and Non-
incendive Equipment for Use in Hazardous Locations

Additional Information:

See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at <http://ul.com/about/locations/>.

